OMRON



- Pain Reliever. Instruction Manual.
- Appareil anti-douleur. Mode d'emploi.
- Schmerztherapiegerät. Gebrauchsanweisung.
- Dispositivo per il trattamento del dolore. Manuale di istruzioni.
- Dispositivo para alivio del dolor. Manual de instrucciones.
- Apparaat ter verlichting van pijn. Gebruiksaanwijzing.
- Электронейромиостимулятор для обезболивания. Руководство по эксплуатации.
- Ağrı Giderici. Kullanım Kılavuzu.




TABLE OF CONTENTS

Before Using the Unit	2
Introduction	
Important Safety Precautions and Warnings	3
Know Your Unit	10
Package Contents	10
Optional Medical Accessories	10
Features	
Buttons and Their Functions	
Operating Instructions	12
Assembly Steps	12
Step 1 – Insert Batteries	12
Step 2 – Attach Electrode Cord to the Main Unit	12
Step 3 – Snap Either Electrode Cord to Each of the Pads	
Step 4 – Remove and Discard Plastic Film from Pads	
Get Started with Your Therapy	14
Step 1 – Pad Placement	
Step 2 – Select 1 of 5 Modes	
Step 3 – Select Intensity Level (1 Low – 10 High)	21

How to Control and Reduce Your Pain	
When should you start therapy?	
Get to your pain early	
How long should you use it?	22
When to stop using the unit?	22
What type of pain is it best for?	22
Cleaning and Storage	23
Cleaning and Storage	
Cleaning the pads	23
When should you replace your pads?	23
Cleaning the unit	
Storing the pads	24
Storing the unit and pads	24
Froubleshooting	25
Warranty	28
Specifications	
mportant information regarding electro magnetic	
Compatibility (EMC)	21

INTRODUCTION

Thank you for purchasing the Omron PocketTens. In order to use the unit safely, read the complete manual carefully before using the unit for the first time. Keep this instruction manual in a convenient place or store with the unit for future reference.

This box contains the following components:

- 1 Main Unit
- 1 Electrode Cord
 2 Long Life Pads
- 1 Pad Holder
- 2 "AAA" Batteries (LR03)
- 1 Instruction Manual

Intended Use

Medical purpose

OMRON PocketTens is a Pain Reliever intended for reducing and relieving muscles and joints pain, stiffness and numbness in the back, arms, legs, shoulders and feet by applying electrical nerve stimulation to the surface of the skin near the site of the pain. It should be applied to normal, healthy, dry, and clean skin of adult patients.

Any of the modes can safely be used on body parts or pains described in this manual. Just find one that feels good and is comfortable on your pain. It can be easily used in conjunction with any other pain treatment or medication.

Since the 1970s, pain relief based on TENS has been widely used by many healthcare professionals, such as physiotherapists and pain specialists.

Intended User

This unit is intended to be operated by adults who can understand this instruction manual. It is intended for home use.

How does it work?

OMRON PocketTens uses Triple Action TENS (Transcutaneous Electrical Nerve Stimulation) technology that helps to:

- block the pain message
- trigger the release of endorphins (natural pain killers)
- improve the blood circulation (as result of repeated muscle contracting and relaxing)

It is important that you read all the warnings and precautions included in this instruction manual because they are intended to keep you safe, prevent injury and avoid a situation that could result in damage to the unit.

SAFETY SYMBOLS USED IN THIS INSTRUCTION MANUAL				
DANGER Improper use may cause danger result death or serious injury. These are situal in which the device should not be used				
	Indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.			
▲ CAUTION	Indicates a potentially hazardous situation which, if not avoided, may result in minor or moderate injury to the user or patient or damage to the equipment or other property.			

Do not use this unit with these other devices:

 If you have a pacemaker, implanted defibrillator, or other implanted metallic or electronic device. Such use could cause electric shock, burns, electrical interference or death.



(2) Do not use this device while using another TENS device.

- (3) Together with a **life-supporting medical electronic device** such as an artificial heart or lung or respirator.
- (4) For Hospitals and Clinics, in the presence of or when attached to the body, electronic monitoring equipment (e.g. cardiac monitors, ECG alarms), which may not operate properly when the electrical stimulation device is in use.
- (5) For Hospitals and Clinics, Simultaneous connection of a patient to a high frequency surgical ME equipment may result in burns at the site of the stimulator electrodes and possible damage to the stimulator.
- (6) For Hospitals and Clinics, operation in close proximity (e.g. 1m) to a shortwave or microwave therapy ME equipment may produce instability to the stimulator output.

DO NOT USE THIS UNIT UNDER THESE CONDITIONS

Consult with your physician before using this unit.

The unit may cause lethal rhythm disturbances in certain susceptible individuals. If you have had a recent surgical procedure, the stimulation may disrupt the healing process.

DO NOT USE ON THESE INDIVIDUALS

A Pregnant women.

- A Do not use on **children or infants** because this device has not been evaluated for pediatric use.
- Keep out of the reach of **young children** because the electrode cord could cause strangulation.
- A Persons incapable of expressing their thoughts or intentions.
- A Persons incapable of operating the unit by themselves.
- A Use caution if you have a tendency to bleed internally, such as following an injury or fracture.

- **If you have suspected or diagnosed heart disease,** you should follow precautions recommended by your physician.
- ▲ If you have suspected or diagnosed epilepsy, you should follow precautions recommended by your physician.
- **A** Use caution if stimulation is applied **over the menstruating uterus**.
- ▲ This device is not intended for use by persons (including children) whose physical, sensory or mental abilities are different or reduced, or lack of experience or knowledge unless such persons are afforded supervision or training to operate the appliance by a person responsible for their safety. Children should be supervised to ensure they do not use the equipment as a toy.

Both sides of the thorax

which could be lethal.

A NEVER APPLY THE PADS TO THESE BODY AREAS:



The **head**, the **mouth**, or any area of the **face**



The **neck** or any area of the **throat** because this could cause severe muscle spasms resulting in closure of the airway, difficulty in breathing, or adverse effects on heart rhythm or blood pressure.



Near the heart, or on genital area.







On the **calves of both legs** at the same time because this may cause cardiac disturbance.

simultaneously (lateral or front and

back), or across your chest because

the introduction of electrical current

may cause rhythm disturbances



On the **soles of both feet** at the same time because this may cause cardiac disturbance.

Open wounds or rashes or over swollen, red, infected or inflamed areas or skin eruptions (such as varicose veins, phlebitis, thrombophlebitis and thrombosis), or on top of or close to cancerous lesions, or over areas of skin that lack normal sensation.

DO NOT USE THIS UNIT DURING THESE ACTIVITIES

- A When in the bath or shower;
- A While sleeping;
- While driving, operating machinery, or during any activity in which electrical stimulation can put you at risk of injury.

PAIN MANAGEMENT WARNINGS

- **A** If you have had medical or physical treatment for your **pain**, **consult** with your physician before using this unit.
- A If your pain does not improve, becomes seriously chronic or severe, or continues for more than five days, stop using the unit and consult with your physician.
- A The mere existence of pain functions as a very important warning telling us that something is wrong. Therefore, if you suffer from any serious illness, consult your physician in order to confirm that it is advisable for you to use this Electronic Nerve Stimulator unit.
- A If you have any infectious disease or illness, consult with your physician before using this unit.

DO NOT ALTER THE UNIT

- A Do not plug the cord into any other device that is not an OMRON PocketTens.
- **A** No modification of this unit is allowed.
- A Use this unit only with the electrode cord and pads recommended by the manufacturer to avoid damage to the unit.

WARNINGS REGARDING THE PADS

- Apply pads to normal, healthy, dry, clean skin (of adult patients) because it may otherwise disrupt the healing process.

- **A** If you experience any skin irritation or redness after a session, do not continue stimulation in that area of the skin.



PRECAUTIONS REGARDING THE PADS

- ▲ Do not move the pads to another location while the unit is on.
 ▲ Therapy won't work with just one pad. You MUST USE TWO PADS at the same time.
- ▲ Make sure the components are connected well and the pads are fixed on the part of the body you wish to treat or the therapy may not be effective.
- Pad should not touch any metal object, such as a belt buckle, necklace, or other metal worn under clothing.
- **Do not overlap pads** or put pads on top of each other. It may weaken or stop therapy, or the unit may stop working.
 - **Gel pads may also stick together** and cause gel to be removed when separating.
- 🗥 Do not apply pads with wet hands.
- ⚠ Do not leave pads attached to the skin after treatment.

- **Do not share pads with another person.** This may cause a skin irritation or infection. Pads are intended for use by one person.
- 🗥 Do not use pads after Exp. date.
- **Do not bend or fold** because the gel may get damaged and it won't stick or function properly.
- ▲ To avoid damage to the adhesive surface of the pads, **put the pads only on the skin or on the plastic pad holder provided.**
- Always place clean pads in accordance with illustrations provided (Refer to pages 14-18, Pad Placement).
- ▲ **Do not apply ointment or any solvent** to the pads or to your skin because it will disrupt the pads from functioning properly. The self-adhesive pads will adhere to your skin.

⚠ Place pads at least 2.5 cm apart for optimal results.

CAUTION WHILE USING UNIT

MAIN UNIT

- ▲ Please report to the manufacturer and the competent authority of the Member State in which you are established about any serious incident that has occurred in relation to this device.
- ▲ If the unit is not functioning properly or you feel discomfort, immediately stop using the unit.
- ▲ **Do not use for any other purpose** except for what it is intended for.
- ▲ Do not place in a room with high humidity, such as a bathroom. This will damage the unit. Ideal temperature for using 10°C 40°C, 30% 80% relative humidity.
- ▲ Do not use the unit without proper lighting. You may not be able to operate unit successfully.
- ▲ While using this device, make sure that no mobile phone or any other electrical devices that emit electromagnetic fields is within 30 cm. This may result in degradation of performance of the device.

CORD

- Do not wash the electrode cord.
- A Do not insert the electrode plug into any place other than the jack

on the unit.

EN

- **Do not pull on the electrode cord** during treatment.
- ightarrow Do not bend or pull the end of the cord.
- \Lambda When pulling out the cord from the unit, **hold the plug and pull.**
- 🗥 Replace the cord when broken or damaged.

BATTERY

- **Do not throw the batteries into a fire.** The batteries may explode.
- Dispose of the unit, batteries, and components according to applicable national regulations. Unlawful disposal may cause environmental pollution.
- **Do not mix alkaline and manganese batteries** as this will shorten the battery life.
- ▲ During therapy, do not remove the battery cover and do not touch the battery terminals.
- \triangle Do not use rechargeable batteries.

PADS

⚠ Use the Pads only recommended by OMRON to avoid affect the

safety and effectiveness of electrical stimulation.

Applying pads incorrectly could result in discomfort or skin burns.
 Detach the pads before replacing the batteries.

General Precautions

- $\underline{\Lambda}$ The long-term effects of electrical stimulation are unknown.
- Apply stimulation to only normal, intact, clean, dry, and healthy skin.
- TENS is not effective in treating the original source or cause of the pain, including headache.
- ▲ TENS is not a substitute for pain medications and other pain management therapies.
- ⚠ TENS devices do not cure disease or injuries.
- ▲ TENS is a symptomatic treatment and, as such, suppresses the sensation of pain that would otherwise serve as a protective mechanism.
- Effectiveness is highly dependent upon patient selection by the guide of a practitioner qualified in the pain management.

- ▲ You may experience skin irritation or hypersensitivity due to the electrical stimulation or electrical conductive medium (gel).
- **Keep unit away from young children.** The unit contains small pieces that may be swallowed. Immediately contact your physician.


Possible Adverse Reactions


- A You should stop using the unit and consult with your physician if you experience adverse reactions from the unit.
- A You may experience skin irritation and burns beneath the stimulation electrodes applied to your skin.
- ▲ Do not use to treat one region for extended periods of time (more than 30 minutes a session, up to 3 times/day) or muscles in that region may become exhausted and sore.

KNOW YOUR UNIT

PACKAGE CONTENTS

Main Unit





Electrode Cord







Pad Holder (1 standard) Batteries* (2 "AAA" batteries (LR03)) * The supplied batteries are for trial use only, they may have a shorter life.

Instruction Manual





OPTIONAL MEDICAL ACCESSORIES

	Product description	Model
	Long Life Pads	HV-LLPAD-E
10		

KNOW YOUR UNIT


FEATURES

- 1. Five pre-set modes (Arm/Shoulder, Lower Back, Leg/Foot, Knead, Steady).
- 2. Ten intensity levels (1 low to 10 high).
- 3. **Pair of pads** (durable, re-usable, washable, up to 150 uses).

Use the pads or cord only recommended by OMRON with this unit.

- 4. Automatic 15 minute shut off.
- 5. **The lights** let you clearly control your therapy with mode, intensity level.
- 6. **Battery Light** lights up to remind you to replace the batteries.
- 7. Pad Light blinks if pad(s) falls off.

BUTTONS AND THEIR FUNCTIONS



ASSEMBLY STEPS

Before using your unit, inspect these things.

- 1. Cord is not broken.
- 2. Pad gel is not damaged.
- 3. Electrode cord connection is not broken.
- 4. Unit is intact and working.
- 5. There is no battery leakage.

STEP 1 – INSERT BATTERIES

1 Remove the battery cover using a hard object (such as the tip of a pen, thin screwdriver).



2 Insert batteries and then install the battery cover. (Make sure the \oplus \ominus signs correspond when inserting batteries).



STEP 2 – ATTACH ELECTRODE CORD TO THE MAIN UNIT

Attach the electrode cord plug to the bottom of the main unit.



ASSEMBLY STEPS

STEP 3 – SNAP EITHER ELECTRODE CORD TO EACH OF THE PADS


For the first time, take the pads out of the sealed package.

 ▲ Do NOT turn unit on, until pads are on your skin.
 ▲ You MUST USE BOTH PADS or stimulation will not work.
 ▲ Don't add spray, lotions or

creams to skin or pads.

NOTE:

Pads will not stick if the skin has too much hair or ointment. Please keep the skin clean. We recommend shaving the area for effective treatment.



For the first time, remove the clear plastic film from the pads and discard it.



(USE FOR A MAXIMUM OF 30 MINUTES PER SESSION)

STEP 1 – PAD PLACEMENT

For optimal therapy:

A Place pads on either side of the pain, not directly on the

pain.

A Place pads at least 2.5 cm **apart**.

MUST USE TWO PADS at a time for therapy to work.

Do not overlap pads or put on top of each other.


▲ Do not share pads.

Before starting your therapy, rate your pain from 1 low to 10 high. This mental check gives you a basis you can compare to once the session is complete.



Attach both pads on either side of the region where you feel pain.

STEP 1 – PAD PLACEMENT (continued)



SHOULDER Attach both pads on the shoulder according to your pain.


SHOULDER


Attach one pad on the front and on the back of your shoulder.



▲ Do not use near the heart, on both sides of the thorax or across your chest because the introduction of electrical current may cause rhythm disturbances which could be lethal.

STEP 1 – PAD PLACEMENT (continued)











LOWER BACK

Attach both pads on the lower back according to your pain. Place pads on muscle of back, not on spine, for optimal therapy. Attach one pad below and above the region in pain, both on same side.

STEP 1 – PAD PLACEMENT (continued)



LEG (HIP & THIGH) Attach both pads on either side of the area with pain.






Do not put the pads on the bottom of both feet at the same time.

STEP 1 – PAD PLACEMENT (continued)



JOINT (ELBOW)

Attach both pads on either side of the joint with the pain.




JOINT (KNEE) Attach both pads above the knee or above and below the joint with pain.

STEP 2 – SELECT 1 OF 5 MODES

① Press the Power button.

- ② Press ▲(Up) or ▼(Down) button to choose 1 of the 5 modes. Modes cannot be combined.
- ③ Press the Set/Start button to select the mode. It will start the therapy at the intensity level of 1.



How to change modes during therapy?

If you want to change modes during therapy, press the Set/Start button and \blacktriangle (Up) or \blacktriangledown (Down) button to select a new mode. You can only use ONE MODE at a time.

If you don't press the Set/Start button to select the mode, the unit will automatically turn off after 3 minutes.

How to select the right mode?

Any of the modes can be used on body parts or pains described in this manual.

Select the mode that feels right for your unique pain.

STEP 2 – SELECT 1 OF 5 MODES (continued)

Therapies designed for	Arm/Shoulder	Lower Back	Leg/Foot	Knead	Steady	EN
Mode Light and the Back of Main Unit	• 1 • 2 • 3 • 4 • 5				● 1 ● 2 ● 3 ● 4 ● 5	
Potential conditions	Stiffness, sore or achy, tight feeling.	Stiffness, soreness, muscle spasm, nerve pain.	Swelling, fatigue, chilly feeling, sore or achy.	Stiffness, soreness, tight feeling.	Stiffness, soreness, tight feeling, or achy.]
What does the therapy deliver?	Series of low to high rate tapping, pulsing, kneading and massage-like sensations.	Series of high rate to low tingling sensations, followed by tapping. With higher intensity, you may feel kneading or massage-like sensations.	Series of low rate tapping, pulsing sensations.	Series of medium rate pulsing sensations to mimic massage.	Series of regular pulsing sensations that do not change.	

STEP 3 - SELECT INTENSITY LEVEL (1 LOW - 10 HIGH)

The unit automatically starts at the intensity level of 1.

Slowly increase the intensity level by pressing \blacktriangle (Up) button. You should feel a gentle pulsing sensation.

How do I select the right intensity level for my pain?


Each time you press \blacktriangle (Up) or \forall (Down) button, it moves to another level. If the stimulation sensation becomes weaker or disappears, increase the intensity. But, if the sensation is at all uncomfortable, press \forall (Down) button to decrease the intensity.

- Press ▲ for higher intensity.
- Press ▼ for lower intensity.

If Pad Light is blinking, the unit will automatically turn off after 30 seconds. ($\ensuremath{\mathbb{CP}}$ "TROUBLE SHOOTING" on page 25)

What intensity level is my unit on?

Press ▲(Up)/▼(Down) button to increase/decrease the intensity. The Intensity Light moves up/down after pressing it once or twice (as shown below). Therefore, the light may not move up/down, but the intensity level does increase/decrease each time you press it.



How long is the therapy?

The unit will continue automatically for 15 minutes before it shuts off. If you want to stop the therapy while in use, press the Power button. We recommend a total of 30 minutes therapy in one sitting, up to 3 times/ day.

When should you start therapy?

Use as soon as your pain begins. Start with one session (the unit automatically turns off at 15 minutes). Turn off with pads still on and RATE YOUR PAIN again (1 low to 10 high).

Get to your pain early

If you get to your pain early, it may prevent the pain from becoming worse, or even chronic. It's better for you to get it under control sooner so that it does not reach a high pain threshold where it limits your daily activities.

How long should you use it?

Start with one 15 minute session. Always turn unit off with pads still on. Rate your pain to check your progress, 1 low to 10 high. Stop therapy session if pain has reduced or stopped. Press the Power button to continue therapy for another 15 minute session.

1 session 15 minutes automatic shut-off	Max minutes/session 2×15 minutes	Max times/day 3 times
---	-------------------------------------	--------------------------

▲ See warnings on page 9. Long-time treatment and strong stimulation may cause muscular fatigue and may generate adverse effects.

When to stop using the unit?

1. If you experienced an **adverse reaction** (skin irritation/ redness/burns, headache or other painful sensation, or if you feel any unusual discomfort).

2. If your pain **does not improve**, becomes seriously chronic and severe, or continues **for more than five days.**

What type of pain is it best for?

This therapy works best on acute pain because it is localized. Acute pain is pain in one area for less than 3 months. If you have chronic pain, you may have pain in more than one area and for longer than 6 months. Chronic pain may be compounded by other issues that this unit cannot address.

Remember, this unit does not cure your pain or the original cause of the pain. It provides temporary relief or reduction of pain so that you can control your life and activities better.

CLEANING AND STORAGE

The unit is designed for repeated use over time. The pads will last up to 150 uses, or 5 months (based on use 1/day). Here are important cleaning and storage instructions:

Cleaning the pads

- 1. Turn the power off and remove the electrode cord from the pads.
- 2. Wash the pads when the adhesive surface becomes dirty and/ or the pads are difficult to attach.
 - Wash the pad softly with your fingertips under slow running cold water for several seconds (do not use a sponge/cloth/ sharp object like a nail onadhesive side, do not use detergents, chemicals or soap).



- 3. Pads can be washed after 15 uses, approximately 10 times for up to 150 uses. Do not wash the pads too long or too frequently. If the adhesive surface becomes sticky or the pad peels off, leave the pad in the refrigerator (do not freeze) overnight. The adhesion may be estored.
- 4. Dry the pads and let the adhesive surface air-dry completely. Do not wipe with a tissue paper or cloth.
- 5. Pads are replaceable and can be purchased. Please consult your local OMRON representatives.

The life of the pads may vary by how often you wash the pads, the skin condition, and how you store the pads.

When should you replace your pads?

If the pad no longer sticks to your skin or if more than 25% of the pad's surface is not in contact with your skin.

CLEANING AND STORAGE

Cleaning the unit

- 1. Turn unit off and disconnect the electrode cord from the pads.
- 2. Clean with a lightly moistened cloth (or a cloth soaked in a neutral cleaning solution) and wipe gently.
 - Do not use chemicals (like thinner, benzene).
 - Do not let water get into the internal area.

Storing the pads

- 1. Turn the unit off and remove the cord from the bottom of the unit.
- 2. Remove the pads from your body.
- 3. Leave the electrode cord connected to the pads.

Place the pads on the pad holder, one pad on each side with the sticky side of each pad on the pad holder.

4. Wrap the electrode cord around the pad holder.



Storing the unit and pads

- Place the unit, pads with electrode cord on pad holder, Instruction Manual inside the original box.
- Do not keep in areas subject to direct sunlight, high or low temperatures, humid area, near to fire, vibration, or shock.

Storage temperature (after unpacking), 0°C - 40°C, 30% - 80% relative humidity.

- Do not keep at places that can be easily reached by children.
- When not in use for a long period, remove the batteries before storage, to avoid liquid discharge from batteries.
- Do not wrap the electrode cords around the unit because it may damage the cord.



TROUBLESHOOTING

In case of any of the below problems occurs during use, first check that no other electrical device is within 30 cm. If the problem persists, refer to the table below.

If this happens	Possible causes	Try this solution	If this happens	Possible causes	Try this solution
The intensity is not felt. Very weak	Are you using only 1 pad?	Put the other pad on your skin. You must use BOTH PADS for therapy to work.	The intensity is not felt.	Are the batteries weak?	Replace both "AAA" batteries.
intensity level.	Have you removed the	Peel off film on the	Very weak intensity level.	Is the intensity "1"?	Press the \blacktriangle (Up) button.
	transparent film from the pads?	adhesive surface of pads.	The skin turns red or the skin feels irritated.	Is the adhesive	Wash adhesive surface
	Are the pads stacked together or do pads overlap?	Check placement of pads. Refer to Pad Placement Guide.		surface of pads dirty or dry?	of pads softly with your fingertips for about 3 seconds under slow running water.
	Is the cord properly connected to the unit?	Connect cord plug correctly into the jack at bottom of this unit.		Is therapy time too long?	Use less than 15 minutes.
	Is the intensity setting getting weak?	Press the ▲ (Up) button.		Are the two pads attached properly to	Refer to the Pad Placement Guide and
	Is the gel damaged?	Replace pad.		the body?	attach correctly.

TROUBLESHOOTING

f this happens	Possible causes	Try this solution	If this happens	Possible causes	Try this solution
The skin turns red or the skin feels irritated.	Is the pad surface worn out?	Replace both pads at the same time.	Pad gel does not stick to skin.	Have you removed the transparent film from the pads?	Peel off film on the adhesive surface of pads.
No power source.	Are the polarities of battery (+ and -) aligned in the wrong direction? or Are the	Check batteries for correct alignment. or Replace batteries.		Is the pad wet? or Is your skin too wet? or Is there any spray, lotions or creams on your skin or pads?	Dry the pad. or Dry the skin.
Power cut off during use.	Are the batteries weak?	Replace both batteries at the same time.		The pad gel may be damaged. Is there too much hair on your skin?	Replace the pad. Shave the immediate area for proper pad
	Is the cord broken?	Replace cord.		Is the adhesive	adhesion. Wash adhesive surface
Battery Light lights up.	Are the batteries weak?	Replace both batteries at the same time.		surface of pads dirty or dry?	of pads softly with you fingertips for about 3 seconds under slow running cold water.

TROUBLESHOOTING

If this happens	Possible causes	Try this solution	
Pad gel does not stick to skin.	Are you using the pads during perspiring?	Dry the pad placement area.	
	Have the pads been washed too long and/ or too frequently?	Leave the pads in refrigerator (do not freeze) for overnight.	
	Were the pads stored under high temperature, high humidity, or direct sunshine?	Replace both pads.	
Pad Light is blinking.	Are both pads attached to the body?	Re-attach dislocated pad(s) onto the skin firmly.	

If this happens	Possible causes	Try this solution
Pad Light is blinking.	Have you removed the transparent film from the pads?	Peel off film on the adhesive surface of pads.
	Is the cord properly connected to the main unit?	Connect cord plug correctly into the jack at the bottom of the main unit.
	Is the adhesive surface of pads dirty or dry?	Wash adhesive surface of pads softly with your fingertips for about 3 seconds under slow running cold water.

If the above measures are not effective, visit our website (www.omron-healthcare.com) for contact information.

WARRANTY

Thank you for buying an OMRON product. This product is constructed of high quality materials and great care has been taken in its manufacturing. It is designed to give you a high level of comfort, provided that it is properly operated and maintained as described in the instruction manual.

This product is warranted by OMRON for a period of 3 years after the date of purchase. The proper construction, workmanship and materials of this product is warranted by OMRON. During this period of warranty OMRON will, without charge for labour or parts, repair or replace the defect product or any defective parts.

The warranty covers only products purchased in Europe, Russia and other CIS countries, Middle East and Africa.

The warranty does not cover any of the following:

- a. Transport costs and risks of transport.
- b. Costs for repairs and / or defects resulting from repairs done by unauthorised persons.
- c. Periodic check-ups and maintenance.

- d. Failure or wear of accessories or other attachments other than the main device itself, unless explicitly warranted above.
- e. Costs arising due to non-acceptance of a claim (those will be charged for).
- f. Damages of any kind including personal caused accidentally or from misuse.

Should (warranty) service be required please apply to the dealer whom the product was purchased from or an authorised OMRON distributor. For the address refer to the product packaging / literature or to your specialised retailer. If you have difficulties in finding OMRON customer services, visit our website (www.omron-healthcare.com) for contact information. Repair or replacement under the warranty does not give rise to any extension or renewal of the warranty period. The warranty will be granted only if the complete product is returned together with the original invoice / cashticket issued to the consumer by the retailer. OMRON reserves the right to refuse the warranty service if any unclear information has been given.

SPECIFICATIONS

Product Category	Electroanalgestic Transcutaneous Stimulators
Product Description	Pain Reliever
Model (code)	PocketTens (HV-F013-E)
Power Source	DC3V (two AAA alkaline batteries or two AAA manganese batteries)
Battery Life	New batteries (two AAA alkaline batteries) will last for approx. 3 months (when used for 15 minutes a day, Lower Back Mode, max. intensity).
Frequency	Approx. 1 to 108Hz
PULSE Duration	100 µsec
Maximum Output Voltage	32V (during 500Ω load)
Power Control	10 intensity levels
Classification of ME equipment	Internally powered
IP classification	IP 22*
Operating Temperature, Humidity, Air Pressure	10 °C to 40 °C, 30 to 80% RH, 700 to 1060 hPa

Storage/Transportation Temperature, Humidity, Air Pressure	-20 °C to 60 °C, 10 to 95% RH, 700 to 1060 hPa
Weight	75g (incl. batteries)
Outer Dimension	Width 75mm × Height 70mm × Depth 22mm
Contents	Main Unit, Electrode Cord, 2 Long Life Pads, Pad Holder, 2 "AAA" Batteries(LR03), Instruction Manual

This OMRON product is produced under the strict quality system of OMRON HEALTHCARE Co., Ltd., Japan.

Designed for a minimum of 5 years life expectancy.

IP22 IP classification is degrees of protection provided by enclosures in accordance with IEC 60529. This device is protected against solid foreign objects of 12.5 mm diameter and greater such as a finger. This device is protected against oblique falling water drops which may cause issues during a normal operation.

NOTE: These specifications are subject to change without notice.

SPECIFICATIONS

C € 0197 Description of symbols that, depending on a model, can be found on the product itself, product sales package or IM			Y	Temperature limitation	
			5	Humidity limitation	
	This product should not be used by persons with medical implants, e.g. heart pacemakers, artificial heart, lung or other electronic life support systems.		$\overline{\langle}$	Atmospheric pressure limitation	
★	Applied part - Type BF Degree of protection against electric shock (leakage current)			Date of Manufacture	
IP22	Ingress protection degree provided by IEC 60529	M	D	Medical Device	
EAC	Symbol of Eurasian Conformity	P	G	GOST-R symbol Serial number	
i	Need for the user to consult the instructions for use	Ictions for use Product production date is integrated in a Serial number, which is placed on the sales package: the first 4 digits mean			
CĒ	CE Marking	yea	year of production, the next 2 digits - month of production.		

EN

IMPORTANT INFORMATION REGARDING ELECTRO MAGNETIC COMPATIBILITY (EMC)

PocketTens manufactured by OMRON HEALTHCARE Co., Ltd. conforms to EN60601-1-2:2015 Electro Magnetic Compatibility (EMC) standard. Further documentation in accordance with this EMC standard is available at OMRON HEALTHCARE EUROPE at the address mentioned in this instruction manual or at www.omron-healthcare.com. Refer to the EMC information for PocketTens on the website.



Correct Disposal of This Product (Waste Electrical & Electronic Equipment)

This marking shown on the product or its literature, indicates that it should not be disposed of, with other household wastes at the end of its working life. To prevent possible harm to the environment or human health from uncontrolled waste disposal, please separate this product from other types of wastes and recycle it responsibly to promote the sustainable reuse of material resources.

Household users should contact either the retailer where they purchased this product, or their local government office, for details of where and how they can return this item for environmentally safe recycling.

Business users should contact their supplier and check the terms and conditions of the purchase contract. This product should not be mixed with other commercial wastes for disposal.

EN

Manufacturer Fabricant Hersteller	Produttore Fabricante Fabrikant	Производитель Üretici الشركة المُصنعة	OMRON HEALTHCARE Co., Ltd. 53, Kunotsubo, Terado-cho, Muko, KYOTO, 617-0002 JAPAN		
EU-representative Mandataire dans l'UE EU-Repräsentant EC REP	Rappresentante per l'UE Представитель в EC Representante en la UE AB temsilcisi Vertegenwoordiging in de EU جهة التمثيل بالاتحاد الأوروبي		OMRON HEALTHCARE EUROPE B.V. Scorpius 33, 2132 LR Hoofddorp, THE NETHERLANDS www.omron-healthcare.com		
Importer in EU Importateur dans l'UE Importeur in der EU	Importatore per l'UE Importador en la UE Importeur in de EU	Импортер в EC AB'de İthalatçı المستورد في الاتحاد الأوروبي			
roduction facility Productiefaciliteit ite de production Производственное подразделение roduktionstätte Üretim Tesisi tabilimento di produzione منشاة (لتصنيع anta de producción			OMRON DALIAN Co., Ltd. No. 3, Song Jiang Road, Economic and Technical Development Zone, Dalian 116600, China		
تلفيرية المعالية الم Sucursales Ounwaha Niederlassungen Yan Kuruluşla Consociate علي المعالية المعالية المعالية المعالية المعالية ال Empresas filiales		ar	OMRON HEALTHCARE UK LTD. Opal Drive, Fox Milne, Milton Keynes, MK15 0DG, UK www.omron-healthcare.com OMRON MEDIZINTECHNIK HANDELSGESELLSCHAFT mbH Konrad-Zuse-Ring 28, 68163 Mannheim, GERMANY www.omron-healthcare.com OMRON Sant E FRANCE SAS 3, Parvis de la Gare, 94130 Nogent-sur-Marne, FRANCE Uniquement pour le marché français: OMRON Service Après Vente N° Vent 0 800 914 31 4 www.omron-healthcare.com		

Made in China Fabriqué en China Hergestellt in China Prodotto in Cina Fabricado en China Geproduceerd in China Cделано в Китае Çin'de Üretilmiştir صنع في الصرين

3271130-2B IM-HV-F013-E-02-02/2020

Issue Date Date de publication Ausgabedatum Data di pubblicación Fecha de publicación : 2020-02-13 Uitgiftedatum Дата banycka Teslim Tarihi تخريض الإصدار